

American National Standard



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

ANSI/AAMI/ ISO 5366-1: 2000

Anaesthetic and
Respiratory Equipment—
Tracheostomy Tubes—
Part 1: Tubes and
Connectors for Use in
Adults



This document was approved and published when the U.S. TAG for TC 121 was held by ASTM, but it is now an AAMI standard. The original formatting has been maintained, so there are some variations from the typical AAMI style.

PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

Published by

Association for the Advancement of Medical Instrumentation
4301 N. Fairfax Drive, Suite 301
Arlington, VA 22203-1633
www.aami.org

© 2014 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of ISO, and AAMI. No part of this publication may be reproduced or distributed in any form, including an electronic retrieval system, without the prior written permission of AAMI. All requests pertaining to this document should be submitted to AAMI. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, complete the reprint request form at www.aami.org or contact AAMI, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633. Phone: 703-525-4890; Fax: 703-525-1067.

Printed in the United States of America

ISBN 1-57020-533-7

**ANS/ISO 5366-1
2000(E)**



Anaesthetic and Respiratory Equipment—

Tracheostomy Tubes—

PREVIEW COPY

Part 1:

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document. For a complete copy of this document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

Tubes and Connectors for Use in Adults

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

Approved as an American National Standard by:

ASTM International



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

These materials are subject to copyright claims of ISO, ANSI, and ASTM International. Not for resale without the prior written permission of ASTM International. No part of this publication may be reproduced in any form, including an electronic retrieval system, without the prior written permission of ASTM International.

INTERNATIONAL STANDARD

ISO 5366-1

Fourth edition
2000-12-15

Corrected and reprinted
2001-09-01



Advancing Safety in Medical Technology

Anaesthetic and respiratory equipment — Tracheostomy tubes —

Part 1: Tubes and connectors for use in adults

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

*Matériel d'anesthésie et de réanimation respiratoire — Tubes de trachéostomie —
Partie 1: Tubes et raccords pour adultes*

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.



Reference number
ISO 5366-1:2000(E)

© ISO 2000

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

© ISO 2000

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.ch
Web www.iso.ch

Printed in Switzerland

Contents

	Page
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Size designation and dimensions	4
5 Materials	6
6 Design and finish	6
7 Requirements for tracheostomy tubes supplied sterile	7
8 Marking and labelling	8

Annexes

A Test method for the security of attachment of connector and neck-plate to tracheostomy tube.....	10
A.1 Principle	10
A.2 Apparatus	10
A.3 Procedure	10
A.4 Expression of results	10
B Test method for determining the resting diameter of the cuff	11
B.1 Principle	11
B.2 Apparatus	11
B.3 Procedure	11
B.4 Expression of results	11
C Guidance on materials and design	12
C.1 Materials	12
C.2 Design	12
Bibliography.....	13

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this part of ISO 5366 may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 5366-1 was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 2, *Tracheal tubes and other equipment*.

This fourth edition cancels and replaces the third edition of ISO 5366-1 and the second edition of ISO 5366-2 (ISO 5366-1:1994 and ISO 5366-2:1993), which have been technically revised.

ISO 5366 consists of the following parts, under the general title *Anaesthetic and respiratory equipment — Tracheostomy tubes*:

- *Part 1: Tubes and connectors for use in adults*
- *Part 3: Paediatric tracheostomy tubes*

Annexes A and B form a normative part of this part of ISO 5366. Annex C is for information only.

Introduction

ISO 5366-1 is one of a series of International Standards dealing with anaesthetic equipment, and is concerned with the basic requirements and method of size designation of tracheostomy tubes made of plastics materials and/or rubber. Specialized tubes, for example those without a connector at the machine end intended for spontaneously breathing patients, and those with reinforced walls or tubes made of metal are excluded from the scope of this part of ISO 5366.

This part of ISO 5366 specifies requirements for tracheostomy tubes with an inside diameter of 6,5 mm or greater. ISO 5366-3 specifies requirements for tracheostomy tubes with an inside diameter from 2,0 to 6,0 mm for paediatric use.

The method of describing tube dimensions and configuration has been devised in order to assist the clinician in the selection of a suitable tube to conform as far as possible to a particular patient's anatomy. Size is designated by inside diameter, which is important because of its relation to resistance to gas flow. Because the stomal and tracheal diameters are important when selecting tubes, it is considered essential that the outside diameter be stated for each size of tube.

Cuffed tracheostomy tubes can be characterized by a combination of the tube inside and outside diameters and by the cuff resting diameter.

The relationship of cuff and tracheal diameters dictates the intra-cuff pressures required to provide a seal. Excessive pressure on the tracheal wall can obstruct capillary blood flow.

A range of cuff designs is available to meet the particular clinical requirements. This part of ISO 5366 requires that the resting diameter of the cuff is marked on the unit package, as this information allows the clinician to match the product to the application.

A 15 mm male conical connector in accordance with ISO 5356-1 should be used for tracheostomy tubes, as for tracheal tubes, to ensure compatibility with the breathing system of an anaesthetic machine or ventilator.

The tracheostomy tube connector should be permanently attached to the tracheostomy tube to prevent inadvertent disconnection of the connector from the tube.

Flammability of tracheostomy tubes, for example if flammable anaesthetics, electrosurgical units, or lasers are used in oxidant-enriched atmospheres, is a well-recognized hazard¹⁾ that is addressed by appropriate clinical management, and is outside the scope of this part of ISO 5366.

1) See ISO/TR 11991.



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

Anaesthetic and respiratory equipment — Tracheostomy tubes —

Part 1:

Tubes and connectors for use in adults

1 Scope

This part of ISO 5366 specifies requirements for tracheostomy tubes made of plastics materials and/or rubber having inside diameters of 6,5 mm or greater. Such tubes are primarily designed for patients who require anaesthesia, artificial ventilation or other respiratory support, but need not be restricted to these uses.

This part of ISO 5366 is not applicable to specialized tubes, and does not address flammability of tracheostomy tubes.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this part of ISO 5366. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this part of ISO 5366 are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 594-1, *Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 1: General requirements.*

ISO 4135, *Anaesthetic and respiratory equipment — Vocabulary.*

ISO 5356-1, *Anaesthetic and respiratory equipment — Conical connectors — Part 1: Cones and sockets.*

ISO 5361, *Anaesthetic and respiratory equipment — Tracheal tubes and connectors.*

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing.*

ISO 11607, *Packaging for terminally sterilized medical devices.*

EN 556 :1994, *Sterilization of medical devices — Requirements for medical devices to be labelled “STERILE”.*

3 Terms and definitions

For the purposes of this part of ISO 5366, the terms and definitions given in ISO 4135 and the following apply.

3.1

tracheostomy tube

tube designed for insertion into the trachea through a tracheostomy

NOTE See Figure 1 a) and b) for an illustration of a typical tracheostomy tube and the associated nomenclature.