

American National Standard



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org

ANSI/AAMI/ ISO 80601- 2-35:2009

Medical electrical
equipment—Part 2-35:
Particular requirements
for the basic safety and
essential performance
of heating devices
using blankets, pads or
mattresses and intended
for heating in medical use



This document was approved and published when the U.S. TAG for TC 121 was held by ASTM, but it is now an AAMI standard. The original formatting has been maintained, so there are some variations from the typical AAMI style.

PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

Published by

Association for the Advancement of Medical Instrumentation
4301 N. Fairfax Drive, Suite 301
Arlington, VA 22203-1633
www.aami.org

© 2014 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of ISO, and AAMI. No part of this publication may be reproduced or distributed in any form, including an electronic retrieval system, without the prior written permission of AAMI. All requests pertaining to this document should be submitted to AAMI. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, complete the reprint request form at www.aami.org or contact AAMI, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633. Phone: 703-525-4890; Fax: 703-525-1067.

Printed in the United States of America

ISBN 1-57020-536-1

INTERNATIONAL
STANDARD

ANSI/IEC
80601-2-35



Edition 2.0
2009-10

Medical electrical equipment —

**Part 2-35: Particular requirements for the
basic safety and essential performance of
heating devices using blankets, pads or
mattresses and intended for heating in
medical use**

This is a preview of this document. It is intended to allow potential purchasers to view the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at
Appareils électromédicaux -
Partie 2-35; Exigences particulières pour la sécurité de base et les
performances essentielles des dispositifs de réchauffage utilisant
des couvertures, des coussins ou des matelas chauffants et
destinés au réchauffage des patients en usage médical

ANSI/IEC 80601-2-35:2009-10



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

These materials are subject to copyright claims of ASTM International. Not for resale without the prior written permission of ASTM International. No part of this publication may be reproduced in any form, including an electronic retrieval system, without the prior written permission of ASTM International.

CONTENTS

FOREWORD.....	v
INTRODUCTION.....	viii
201.1 Scope, object and related standards	1
201.2 Normative references	3
201.3 Terms and definitions	3
201.4 General requirements.....	6
201.5 General requirements for testing ME EQUIPMENT	7
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	7
201.7 ME EQUIPMENT identification, marking and documents.....	7
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	11
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	16
201.10 Protection against unwanted and excessive radiation HAZARDS	18
201.11 Protection against excessive temperatures and other HAZARDS	18
201.12 Accuracy of controls and instruments and protection against hazardous outputs.....	21
201.13 HAZARDOUS SITUATIONS and fault conditions.....	26
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	31
201.15 Construction of ME EQUIPMENT.....	31
201.16 ME SYSTEMS.....	36
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	36
202 Electromagnetic compatibility – Requirements and tests	36
208 General requirements, tests and guidance for ALARM SYSTEMS in MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS	36
210 * Requirements for the development of physiologic closed-loop controllers	37
Annex D (informative) Symbols on marking.....	37
Annex AA (informative) Particular guidance and rationale	38
Annex BB (normative) Determination of the LAGGING MATERIAL	50
Annex CC (normative) *Determination of heat transfer towards the PATIENT	51
Annex DD (normative) *Determination of heat transfer away from the PATIENT	53
Annex EE (normative) Conditions of adequate heat discharge.....	54
Annex FF (normative) Test procedure for maximum CONTACT SURFACE TEMPERATURE for FORCED AIR DEVICES	55
Annex GG (normative) Test procedure for maximum CONTACT SURFACE TEMPERATURE for FORCED AIR DEVICES under SINGLE FAULT CONDITION.....	57
Annex HH (normative) Safety test procedure for average CONTACT SURFACE TEMPERATURE for FORCED AIR DEVICES	58
Bibliography.....	60
Index of defined terms used in this particular standard.....	61

Figure 201.101 – Positioning of temperature sensors on the contact surface of the heated area of a HEATING DEVICE (see 201.12.4.101 and 201.12.4.105)12

ANSI/IEC 80601-2-35:2009

Figure 201.102 – Example of the positioning of temperature sensors on the contact surface of the heated areas of a HEATING DEVICE having more than one separately heated area	5
Figure 201.103 a) – Apparatus for the spark ignition test – Detail A: The apparatus (see 201.8.8.4.101).....	13
Figure 201.103 b) – Apparatus for the spark ignition test – Detail B: Lower member of mask	14
Figure 201.103 c) – Apparatus for the spark ignition test – Detail C: Upper member of mask	14
Figure 201.103 – Apparatus for the spark ignition test	14
Figure 201.104 – Ramp for the impact test on PADS	17
Figure 201.105 – Partial covering conditions	19
Figure 201.106 – Method of folding BLANKETS	28
Figure 201.107 – Examples of folds.....	30
Figure 201.108 – Positions of a BLANKET for the RUCK-RESISTANCE test	35
Figure HH.1 – Sensor locations – Average CONTACT SURFACE TEMPERATURE	59
Table 201.101 – *Additional ESSENTIAL PERFORMANCE requirements.....	7
Table 201.102 – Temperature limits in dependency to time	33

PREVIEW COPY
This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

**Part 2-35: Particular requirements for the basic safety
and essential performance of heating devices using blankets,
pads or mattresses and intended for heating in medical use**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 80601-2-35 has been prepared by IEC technical committee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice, and subcommittee 1: Breathing attachments and anaesthetic machines, of ISO technical committee 121: Anaesthetic and respiratory equipment.

This second edition cancels and replaces the first edition published in 1996. This edition constitutes a technical revision.

This new edition provides consistency with the third edition of IEC 60601-1, as well as with the four other particular standards related to paediatric equipment for which the committee is responsible.

ANSI/IEC 80601-2-35:2009

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62D/784A/FDIS	62D/804/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table. In ISO, the standard has been approved by 15 P-members out of 15 having cast a vote.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment* can be found on the IEC website.

The committee has decided that the contents of this particular standard will remain unchanged until the maintenance result date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.



IMPORTANT – The “colour inside” logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this publication using a colour printer.

PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

ANSI/IEC 80601-2-35:2009

INTRODUCTION

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation for heating devices using BLANKETS, PADS or MATTRESSES and intended for heating in medical use.

This particular standard amends and supplements IEC 60601-1 (third edition, 2005) *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*, hereinafter referred to as the general standard. The text of this particular standard relating to forced air warmers is based on ASTM F2196-02, *Standard specification for circulating liquid and forced air patient temperature management devices*.

The requirements are followed by specifications for the relevant tests.

A "general guidance and rationale" section giving some explanatory notes, where appropriate, about the more important requirements is included in Annex AA.

While K (degree Kelvin) is the recognized unit and symbol for absolute temperature and temperature difference, °C has been used throughout this particular standard because all measurements are commonly made using equipment marked with the Celsius temperature scale.

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-35: Particular requirements for the basic safety and essential performance of heating devices using blankets, pads or mattresses and intended for heating in medical use

201.1 Scope, object and related standards

Clause 1 of the general standard¹⁾ applies, except as follows:

201.1.1 * Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of HEATING DEVICES using BLANKETS, PADS or MATTRESSES in medical use, also referred to as ME EQUIPMENT. HEATING DEVICES intended to prewarm a bed are included in the scope of this International Standard.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

If a clause or subclause is specifically intended to apply to a specifically defined type of ME EQUIPMENT, as is the case with FORCED AIR DEVICES, then the clause or subclause is entitled as such. Clauses or subclauses that apply to all types of ME EQUIPMENT within the scope of this standard are not specifically entitled.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard.

This particular standard does not apply to:

- HEATING DEVICES intended for physiotherapy;
- radiant warmers; for information, see IEC 60601-2-21 [12]²⁾;
- incubators; for information, see IEC 60601-2-19 [10];
- transport incubators, for information, see IEC 60601-2-20 [11];
- cooling devices.

201.1.2 Object

Replacement:

¹⁾ The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

²⁾ Figures in square brackets refer to the Bibliography.