

# Technical Information Report

PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit [www.aami.org](http://www.aami.org).

## AAMI TIR61: 2014

Generating reports for  
human factors design  
validation results  
for external cardiac  
defibrillators



## PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit [www.aami.org](http://www.aami.org).



## PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

# Generating reports for human factors design validation results for external cardiac defibrillators

+1-877-249-8226 or visit [www.aami.org](http://www.aami.org).

Approved 24 October 2014 by  
**Association for the Advancement of Medical Instrumentation**

**Abstract:** Provides guidance on the formatting and content of reports generated for the purpose of submitting human factors data for evaluation.

**Keywords:** defibrillation, human factors, usability, external defibrillator, automated external defibrillator

## AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved, usually every five years but at least every ten years. For a TIR, AAMI consults with a technical committee about five years after the publication date (and periodically thereafter) for guidance on whether the document is still useful; i.e., to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice, or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

**CAUTION NOTICE:** This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field of technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633.

AAMI acknowledges the contributions of the AAMI Human Factors group in the development of this TIR.

*Published by*

Association for the Advancement of Medical Instrumentation  
4301 N. Fairfax Dr., Suite 301  
Arlington, VA 22203-1633  
[www.aami.org](http://www.aami.org)

© 2014 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, complete the reprint request form at [www.aami.org](http://www.aami.org) or contact AAMI at 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633. Phone: +1-703-525-4890; Fax: +1-703-525-1067.

Printed in the United States of America

**ISBN 1-57020-568-x**

# Contents

Page

Committee representation.....	iv
Foreword.....	v
Introduction.....	vi
<b>1</b> Scope.....	1
<b>2</b> Object.....	1
<b>3</b> Definitions.....	1
<b>3.1</b> Human factors (HF) engineering.....	1
<b>3.2</b> Usability.....	1
<b>3.3</b> User-centered design (UCD).....	1
<b>3.4</b> User interface (UI).....	1
<b>3.5</b> Human Factors Validation testing.....	2
<b>4</b> Validation study.....	2
<b>5</b> Contents of a validation report.....	2
<b>5.1</b> Format.....	2
<b>5.2</b> Executive summary.....	3
<b>5.3</b> Purpose.....	3
<b>5.4</b> Methods.....	3
<b>5.5</b> Results.....	6
<b>5.6</b> Deviations and exclusions.....	7
<b>5.7</b> Discussion.....	7
<b>5.8</b> Conclusions.....	7
<b>6</b> References.....	8
<b>Annex:</b> Design Considerations and Best Practices.....	9



**PREVIEW COPY**

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit [www.aami.org](http://www.aami.org).

## Committee representation

### Association for the Advancement of Medical Instrumentation

#### AAMI/DF Defibrillator Committee

This Technical Information Report was developed by the AAMI/DF Defibrillator Committee. The draft of this document was circulated to the AAMI/HE Human Factors Engineering Committee for comments. Committee approval of the Technical Information Report does not necessarily imply that all committee members voted for its approval.

At the time this document was published, the **AAMI/DF Defibrillator Committee** had the following members:

- Chairs:* Oscar Tovar-Calderon, MD, FDA/CDRH  
Robert J. Zito, Physio-Control
- Members:* Joseph Basta, Spacelabs Medical Inc.  
Mohamud Daya, MD, MS, Oregon Health & Science University  
Regis DeSilva, MD, Beth Israel Deaconess Medical Center  
Sreeram Dhurjaty, Dhurjaty Electronics Consulting LLC  
Wally R. Elliott, CCE, University of Vermont  
David K. Hunt, Philips Electronics North America  
Nicole Hurley, PhD, WL Gore & Associates Inc  
Janice M. Jenkins, PhD, University of Michigan School of Dentistry  
Amanda Jones, Health Canada  
Carolyn Lall, Draeger Medical Systems Inc.  
Dongping Lin, PhD  
Shen Luo, Mindray DS USA Inc.  
Bokang C. Rapoo, MS, New York Presbyterian Hospital  
David M. Selvitelli, Covidien  
Willis A. Tacker, Jr., MD, PhD, Purdue University  
Kok-Swang Tan, PhD, Medical Devices Bureau Health Canada  
Jeffrey Wiser, 3M Healthcare  
Brian J. Young, GE Healthcare
- Alternates:* Fred W. Chapman, PhD, Physio-Control  
David Hengl, Draeger Medical Systems Inc  
Jayna Stohlman, FDA/CDRH  
Ann Thorn, Philips Electronics North America  
Michael J. Walt, 3M Healthcare

---

NOTE—Participation by federal agency representatives in the development of this technical information report does not constitute endorsement by the federal government or any of its agencies.

---

## Foreword

This technical information report was developed by the AAMI/DF, Defibrillator Committee. The draft of this document was circulated to the AAMI/HE, Human Factors Engineering Committee for comments. The objective is to provide guidance on the Human Factors Engineering (HFE) and usability for external cardiac defibrillators.

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every five years but at least every 10 years. For a TIR, AAMI consults with a technical committee about five years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice, or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

**CAUTION NOTICE:** This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field of technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633.

---

NOTE—This foreword does not contain provisions of AAMI TIR61:2014, *Generating reports for human factors design validation results for external cardiac defibrillators*, but it does provide important information about the development and intended use of the document.

---

## Introduction

External cardiac defibrillators and defibrillator/monitors (defibrillators) treat the most common cause of sudden cardiac arrest, called ventricular fibrillation (VF) and certain ventricular tachycardias by providing the only effective treatment for survival of sudden cardiac arrest (SCA): electrical energy (i.e., a shock) to the heart in order to interrupt the arrhythmia and allow the heart to reestablish a normal rhythm. Defibrillators are indicated for use on unconscious, non-responsive victims who have collapsed due to suspected SCA. The user populations for defibrillators range from untrained lay users to emergency medical services (EMS) personnel to highly trained medical professionals. Use patterns also vary widely. Defibrillators can be found in private homes, public access defibrillator (PAD) programs, corporate and industrial settings, airports, schools, emergency medical response and police vehicles, and hospitals.

Defibrillators are used in urgent, time-critical conditions. The operators, especially home users, are likely to be stressed when they must use the automated external defibrillator (AED) to save a life of a patient or family member. Defibrillators deliver a significant electrical shock and therefore must be safe for the patient bystander and the user. Some defibrillators provide automated or semi-automated defibrillation only; others may provide ECG monitoring and CPR coaching with a feedback device. Still others include additional functions, such as SpO<sub>2</sub>, 12-lead ECG, and capnography. The complexity of the defibrillator determines the extent and type of user testing and validation activities required.

Due to the wide range of training and medical knowledge of the potential users of defibrillators, data-driven design of the user interface is key to minimizing use errors.

To achieve a successful interaction design solution, the defibrillator industry should employ human factors (HF)/user-centered design (UCD) process of the proposed defibrillator system.

The design of the defibrillator should minimize the likelihood of use-related failures. The differences in user type and use environments should be evaluated and addressed in the user interface design process. For example, if the intended user of the defibrillator is a home user, then that user might be most effectively guided by the defibrillator's functional design rather than by printed labeling (e.g., instruction manuals). The defibrillator's user interface and sequences of operation (such as voice prompts, primary controls, indicators, and labels) should consider if the users should be able to operate it without training. Further, it is recommended that, prior to validating the usability of the final design through testing, the design process should be guided through formative evaluations to ensure that by the time Human Factors validation testing is done, the design is safe and effective for its intended users and use environments.

This TIR highlights the recommended content of human factors usability validation reports. Appendix A includes recommendations and considerations for the human factors design and formative design evaluation of cardiac defibrillators.

# Generating reports for human factors design validation results for external cardiac defibrillators

## 1 Scope

This Technical Information Report (TIR) applies to the human factors design of external cardiac defibrillators, as covered by ANSI/AAMI/IEC 60601-2-4:2010, *Medical electrical equipment – Part 2-4: Particular requirements for the basic safety and essential performance of cardiac defibrillators*. The guidance provided by this TIR is intended to be applicable to all external cardiac defibrillator/monitors, and automated external defibrillators (AEDs) including public access defibrillators (PADs).

## 2 Object

The object of this TIR is to provide guidance on the content of human factors validation reports for external cardiac defibrillators. The intention of providing this guidance is to ensure that the submitted report includes (or, depending on design complexity of the defibrillator, reports include) sufficient and consistent human factors validation data for external cardiac defibrillators.

Appendix A provides considerations for the formative stages of product design and development to achieve successful design and validation of defibrillators.

This is a preview of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit [www.aami.org](http://www.aami.org).