

Technical Information Report

PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

AAMI TIR34: 2014

Water for the reprocessing
of medical devices



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.



Water for the reprocessing of medical devices

PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

Approved 4 August 2014 by
Association for the Advancement of Medical Instrumentation

For a complete copy of this AAMI document, contact AAMI at
1-877-749-8226 or visit www.aami.org.

Abstract: This technical information report (TIR) covers the selection and maintenance of effective water quality suitable for reprocessing medical devices. It provides guidelines for selecting the water quality necessary for the reprocessing of categories of medical devices and addresses water treatment equipment, water distribution and storage, quality control procedures for monitoring water quality, strategies for bacterial control, and environmental and personnel considerations.

Keywords: carbon filters, deionization, disinfection, distillation, pasteurization, reverse osmosis, sediment filters, sterilization, ultrafiltration, water filtration, water quality, water softening, water treatment

AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every 5 years but at least every 10 years. For a TIR, AAMI consults with a technical committee about 5 years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice, or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

CAUTION NOTICE: This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 4301 N. Fairfax Dr., Ste. 301, Arlington, VA 22203-1633.

Published by

Association for the Advancement of Medical Instrumentation
4301 N. Fairfax Dr., Ste. 301
Arlington, VA 22203-1633
www.aami.org

© 2014 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact AAMI at 4301 N. Fairfax Dr., Ste. 301, Arlington, VA 22203-1633. Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

ISBN 1-57020-544-2

Contents

Page

Glossary of equivalent standards.....	v
Committee representation.....	vii
Foreword.....	viii
Introduction.....	1
1. Scope.....	2
1.1 General.....	2
1.2 Inclusions.....	2
1.3 Exclusions.....	2
2. Definitions and abbreviations.....	2
3. The importance of water quality and effective water treatment.....	8
3.1 Introduction.....	8
3.2 Major impacts of adverse water quality on medical device reprocessing.....	8
3.2.1 General considerations.....	8
3.2.2 Microbial level in water.....	8
3.2.3 Inorganic and organic components of water.....	9
3.2.4 Water temperature.....	10
3.3 Treatment of water.....	10
3.3.1 General considerations.....	10
3.3.2 Pretreatment.....	11
3.3.3 Principal water treatment processes.....	11
3.3.4 Distribution.....	11
3.4 Categories of medical devices.....	12
3.5 Stages of medical device reprocessing in which water quality is a consideration.....	13
4. Categories of water quality for medical device reprocessing.....	15
4.1 Introduction.....	15
4.2 Two categories of water quality.....	15
5. Selection of water quality.....	18
5.1 Introduction.....	18
5.2 Cleaning.....	18
5.2.1 Manual cleaning.....	18
5.2.2 Mechanical cleaning by medical washers and medical washer–disinfectors.....	18
5.2.3 Automated cleaning by ultrasonic cleaners.....	20
5.3 Disinfection and sterilization.....	20
5.3.1 General considerations.....	20
5.3.2 Medical devices that receive steam sterilization or low-temperature gas sterilization.....	21
5.3.3 Medical devices that receive liquid chemical high-level disinfection.....	23
5.3.4 Medical devices that receive liquid chemical sterilization.....	25
5.3.5 Medical devices that receive pasteurization or thermal disinfection.....	27
6. Water treatment systems.....	28
6.1 Introduction.....	28
6.2 General issues associated with water treatment.....	28
6.3 Design of water treatment systems.....	29
6.3.1 General considerations.....	29
6.3.2 Physical layout of the water purification, distribution, and storage system.....	29
7. Monitoring water quality.....	30
7.1 Introduction.....	30
7.2 Goals of water quality monitoring.....	30
7.3 Water characteristics that should be monitored.....	30
7.3.1 General considerations.....	30
7.3.2 Water temperature.....	30

8	Strategies for bacterial control	33
9	Personnel considerations.....	33
9.1	Introduction.....	33
9.2	Device reprocessing personnel.....	33
9.3	Water maintenance personnel.....	33
9.4	Audits.....	33
10	Continuous quality improvement.....	33
10.1	Introduction.....	33
10.2	Quality process.....	33

Annexes

A	Water treatment methods	35
B	Monitoring water treatment equipment and processes	45
C	Water storage and distribution.....	51
D	Strategies for bacterial control	53
E	Thermal disinfection.....	57
F	Water treatment using filtration.....	58
G	Typical presentation of water quality issues during the reprocessing of medical devices.....	60
H	Bibliography.....	63

Tables

1	Categories and recommended levels of water quality for medical device reprocessing	16
2	Water quality for reprocessing devices to be sterilized by steam or low-temperature gas.....	22
3	Water quality for reprocessing devices to be high-level disinfected.....	24
4	Water quality for reprocessing devices to receive liquid chemical sterilization	26
5	Water quality for reprocessing devices to be pasteurized or thermally disinfected.....	27
6	Overview of water quality monitoring.....	32
7	Quality monitoring of cleaning, disinfection, and sterilization equipment	34
A.1	Summary of water treatment methods.....	40
A.2	Processes that can remove interfering compounds that might be found in water.....	41
B.1	Monitoring water treatment equipment.....	47
E.1	Holding times in instrument washers	57
G.1	Examples of observed problems during device reprocessing that can be caused by poor water quality	60

Figures

1	Stages of medical device reprocessing in which water quality is a consideration.....	14
2	Example of a recommended general water treatment process for incoming water to produce treated water that is appropriate for use in medical device reprocessing.....	28
A.1	Examples of water treatment processes to produce critical and softened water.....	42
A.2	Example of a water treatment process	42
A.3	Example of a water treatment process	42
A.4	Example of a water treatment process	42

Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. AAMI maintains a current list of each International Standard that has been adopted by AAMI (and ANSI). Available on the AAMI website at the address below, this list gives the corresponding U.S. designation and level of equivalency to the International Standard.

www.aami.org/standards/glossary.pdf



PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

Committee representation

Association for the Advancement of Medical Instrumentation

AAMI Water Quality for Medical Device Reprocessing Working Group

This technical information report (TIR) was developed by the AAMI Water Quality for Medical Device Reprocessing Working Group under the auspices of the AAMI Sterilization Standards Committee. Approval of the TIR does not necessarily mean that all working group members voted for its approval. At the time this TIR was published, the Water Quality for Medical Device Reprocessing Working Group had the following members:

- Co-chairs:* Jacqueline Daley, Sinai Hospital of Baltimore
Emily Mitzel, MS, Nelson Laboratories Inc
- Members:* James Weldon Baker, AmeriWater
Ralph J. Basile, MBA, Healthmark Industries Company Inc
Nola Bayes, MBA, Sanford Health
Chris Bible, Belimed Inc
David J. Brodersen, Covidien
Jennifer Burrell, Integrated Medical Systems
Harriet Chan-Myers, Johnson & Johnson
Nancy Chobin, RN CSPDM, St Barnabas Healthcare System
Matthew Dehmler, Bausch & Lomb Inc
Stacy DeMoss, NAMSA
Mary Ann Drosnock, MS, Olympus America Inc
Betty D. Edge, North Shore University Hospital
Marcia Ann Frieze, Case Medical Inc
Steve N. Goldstine, PhD, Steve Goldstine Consultants
Shelley Green, WuXi AppTec Inc
Charles Ofen Hancock, RAC, H&W Technology LLC
Douglas E. Harbrecht, Sterility Assurance LLC
Sandra Iverson, Medivators Inc
Nupur Jain, Intuitive 0053urgical Inc
Nancy E. Kaiser, Steris Corporation
Md Sajedul Islam Khan, II, Apollo Hospital Dhaka
Carolyn L. Kinsley, LexaMed Ltd
Susan G. Klacik, CCSMC FCS ACE, IAHCSSM
Colleen Patricia Landers, RN, Timmins & District Hospital
Jo Ann Barbara Maltais, PhD, Maltais Consulting
Teckla A. Maresca, LPN CSPDM, St Clare's Health System
Rodney D. Parker, Stryker Instruments Division
Patrick Polito, Moog Medical Devices
Linda Slone, RN BSPA CNOR
Donald Socha, Jr., Getinge USA
Joan M. Spear, MBA RM CMPR, B Braun of America Inc
Karen Swanson, Connecticut Childrens Medical Center
Sharon Van Wicklin, MSN RN CNOR/CRNFA, Association of Perioperative Registered Nurses
Nora E. Wikander, RN, CSPDM, St Josephs Wayne Hospital
Kelvin J. Witcher, 3M Healthcare
Martha L. Young, Martha L Young, LLC
- Alternates:* Denise Adams, B Braun of America Inc
Mike Cain, Getinge USA
Claudia Camp, Stryker Instruments Division
Charles Cogdill, Covidien
Ramona Conner, RN MSN CNOR, Association of Perioperative Registered Nurses
Robert Dumont, Case Medical Inc
Gordon M. Ely, WuXi AppTec Inc
Brent Geiger, MS RAC, Medivators Inc
Stephen M. Kovach, Healthmark Industries Company Inc
Chris Lau, LexaMed Ltd
Seth Masek, Covidien
Gerald E. McDonnell, PhD, Steris Corporation
Joelle Pelletier, TSO3 Inc
Joy Perry, Moog Medical Devices

Michelle Pierce, NAMSA
Michael Quin, Johnson & Johnson
Michael J. Schoene, Bausch & Lomb Inc
Grace A. Thornhill, PhD, 3M Healthcare
Sara Toole, BS RM, Nelson Laboratories Inc
Donna Ungvarsky, MSN MEd RN CNOR, Olympus America Inc
Brian Wallace, Intuitive Surgical Inc
Jonathan A. Wilder, PhD, H&W Technology LLC

NOTE—Participation by federal agency representatives in the development of this document does not constitute endorsement by the federal government or any of its agencies.

At the time this document was published, the **AAMI Sterilization Standards Committee** had the following members:

Co-chairs: Victoria M. Hitchins, PhD, FDA/CDRH
Michael H. Scholla, Dupont Protection Technologies

Members: Christopher Anderson, Boston Scientific Corporation
Trabue D. Bryans, BryKor LLC
Nancy Chobin, RN CSPDM, St Barnabas Healthcare System
Charles Cogdill, Covidien
Ramona Conner, RN MSN CNOR, Association of Perioperative Registered Nurses
Jacqueline Daley, Sinai Hospital of Baltimore
Kimbrell Darnell, CR Bard
Lisa Foster, Medpoint LLC
Joel R. Gorski, PhD, NAMSA
Joyce M. Hansen, Johnson & Johnson
Douglas F. Harbrecht, Sterility Assurance LLC
Deborah A. Havlik, Hospira Worldwide Inc
Susan G. Klacik, CCSMC FCS ACE, IAHC SMM
Byron J. Lambert, PhD, Abbott Laboratories
Colleen Patricia Landers, RN, Timmins & District Hospital
Reynaldo Lopez, Cardinal Health (MP&S)
Jeffrey Martin, Alcon Laboratories Inc
Patrick J. McCormick, PhD, Bausch & Lomb Inc
Gerald E. McDonnell, PhD, Steris Corporation
Janet M. Prust, 3M Healthcare
Nancy Rakiewicz, Moog Medical Devices
Mark Seybold, Baxter Healthcare Corporation
Andrew Sharavara, PhD, Propper Manufacturing Co Inc
Mark N. Smith, Getinge USA
James Sidney Wiggs, BSN CRCST, Legacy Health System
Martell Kress Winters, BS SM, Nelson Laboratories Inc
William E. Young
William T. Young, Sterigenics International

Alternates: Lloyd Brown, Covidien
Peter A. Burke, PhD, Steris Corporation
Dave Dion, Cardinal Health (MP&S)
Ken Eddington, NAMSA
Gordon M. Ely, WuXi AppTec Inc
Thomas J. Frazar, Johnson & Johnson
Martha M. Kadas, Sterigenics International
Jim Kaiser, Bausch & Lomb Inc
Natalie Lind, IAHC SMM
Ralph Makinen, Boston Scientific Corporation
Mary S. Mayo, CR Bard
David Ford McGoldrick, BS, Abbott Laboratories
Jerry R. Nelson, PhD, Nelson Laboratories Inc
Patrick Polito, Moog Medical Devices
Karen Polkinghorne, Dupont Protection Technologies
Mike Sadowski, Baxter Healthcare Corporation
Craig A. Wallace, 3M Healthcare

NOTE—Participation by federal agency representatives in the development of this document does not constitute endorsement by the federal government or any of its agencies.

Foreword

This technical information report was developed by the AAMI Water Quality for Medical Device Reprocessing Working Group under the auspices of the AAMI Sterilization Standards Committee. The objective of this TIR is to provide guidelines to personnel involved in medical device reprocessing on the quality of water that should be used in various stages of medical device reprocessing. It is also intended to provide guidelines to water service maintenance personnel on establishing and monitoring water treatment systems. This second edition has been changed to reflect two water categories instead of the four previously specified. There are also many updates and organizational enhancements.

The concepts incorporated in this TIR should not be considered inflexible or static. The recommendations presented here will be reviewed and updated periodically to reflect new information and technical developments regarding water quality and treatment in medical device reprocessing.

Suggestions for improving this TIR are invited. Comments and suggested revisions should be sent to Technical Programs, AAMI, 4301 N. Fairfax Dr., Ste. 301, Arlington, VA 22203-1633.


Advancing Safety in Medical Technology

PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at +1-877-249-8226 or visit www.aami.org.

Introduction

Water quality is an important consideration in all stages of medical device reprocessing. Ensuring adequate water quality in device reprocessing requires collaboration between the personnel who reprocess medical devices and the personnel who establish and maintain the water treatment system. Because the needs of these two groups are distinct, this technical information report (TIR) contains:

- a) sections in the main text that provide guidance for personnel involved in medical device reprocessing on the selection of the recommended water quality for each stage of medical device reprocessing for each category of medical device; and
- b) annexes that provide technical information to water maintenance personnel (i.e., personnel who are involved in water treatment and distribution in the facility) to guide them in setting up and monitoring water treatment systems.

Water can be treated by a variety of methods that yield different levels of water quality. In general, as the chemical quality of water improves, its microbial content could increase unless the system is closely monitored to prevent microbial overgrowth. Gram-negative bacteria and nontuberculous mycobacteria can grow in any type of water, including tap, softened, deionized (DI), reverse osmosis (RO) treated, and distilled water. The rate of growth and the microbial levels attained are a function of the amount of organic contaminants in the water. The importance of monitoring water quality to prevent problems with microbial overgrowth cannot be overemphasized.

This TIR defines two levels of water quality suitable for medical device reprocessing, and it describes the water treatment processes that can be used to obtain the correct water quality. To provide optimal water for medical device reprocessing, reprocessing personnel and water maintenance personnel should collaborate with administrative personnel to implement the following procedures:

This is a preview edition of an AAMI guidance document and is

Step	Procedure	What to Do	Who is Responsible
1	Assessment of water quality	The tap water from the public utility source should be analyzed by an accredited facility with expertise in water quality to determine whether the water requires treatment and, if so, what type of treatment. This analysis should take into account seasonal variations in water quality.	Water maintenance personnel
2	Implementation of water treatment process	On the basis of the assessment in Step 1 and in consultation with an accredited facility with expertise in water quality, personnel should ensure that treatment processes are implemented to provide the type of water quality needed for the medical device reprocessing needs of the facility.	Water maintenance personnel in conjunction with device reprocessing personnel
3	Assurance of proper water quality for the various stages in medical device reprocessing	Medical device reprocessing areas should be audited to determine whether water of the correct quality is being used for the devices being reprocessed in each area. If not, the water treatment should be modified as necessary.	Device reprocessing personnel in conjunction with health care technology management personnel
4	Ongoing monitoring of water quality	Where applicable, monitoring procedures should be established to ensure that the treated water is of adequate quality for medical device reprocessing. Water maintenance personnel and device reprocessing personnel should communicate effectively to ensure that action is taken when inadequate water quality is detected.	Water maintenance personnel in conjunction with device reprocessing personnel

This TIR was developed using information on the reprocessing of medical devices labeled for reuse. There might be regulatory requirements or other issues associated with reprocessing single-use devices that were not considered in developing this TIR (see www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/ReprocessingofSingle-UseDevices/default.htm).

Water for the reprocessing of medical devices

1. Scope

1.1 General

This TIR addresses how to determine the water quality needs for reprocessing various categories of medical devices at various stages of reprocessing and how to assess, generate, monitor, and maintain water meeting those requirements.

1.2 Inclusions

This TIR covers the quality of the water used to clean, rinse, disinfect, and sterilize medical devices. It defines water types on the basis of hardness, pH, microorganism levels, endotoxin levels, and other characteristics. The following specific topics are covered:

- a) Importance of water quality and effective water treatment
- b) Categories of water quality for medical device reprocessing
- c) Selection of water quality
- d) Water treatment systems
- e) Monitoring of water quality
- f) Strategies for microorganism control
- g) Personnel considerations
- h) Continuous quality improvement
- i) Troubleshooting water quality issues

This TIR also provides definitions of terms and a bibliography. The annexes contain technical details pertaining to water treatment and monitoring for the benefit of water maintenance personnel.

1.3 Exclusions

This TIR does not cover the water requirements for hemodialysis applications. See ANSI/AAMI/ISO 13959, ANSI/AAMI/ISO 11663, and ANSI/AAMI RD47.

This TIR does not address water treatment performed within medical washers, washer–disinfectors, or automated endoscope reprocessors (AERs); it covers only the water coming into such equipment. The internal filtering or other additional treatment performed within a medical washer, washer–disinfector, or AER is the responsibility of the equipment manufacturer. However, it is in the best interest of reprocessing personnel to be informed of the internal water treatment apparatus and of the equipment manufacturer to take into account the water quality recommendations of this document. See the ANSI/AAMI and ISO 15883 series of documents.