

# Technical Information Report

AAMI TIR42:2010



This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document,  
contact AAMI at (877) 249-8226  
or visit [www.aami.org](http://www.aami.org).

## **Evaluation of particulates associated with vascular medical devices**



Association for the Advancement  
of Medical Instrumentation



This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document,  
contact AAMI at (877) 249-8226  
or visit [www.aami.org](http://www.aami.org).



This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document,  
contact AAMI at (877) 249-8226  
or visit [www.aami.org](http://www.aami.org).

## **Evaluation of particulates associated with vascular medical devices**

Approved 13 December 2010 by  
**Association for the Advancement of Medical Instrumentation**

**Abstract:** This document addresses vascular exposure to particles arising from the manufacturing environment for medical devices and from the use of medical devices. This TIR is intended to offer guidance to medical device manufacturers in applying analytical methods for particulate testing, identifying potential sources of particulates, and developing limits for particulates.

**Keywords:** component, control, contaminant, contamination, environment, evaluation, device, identity, limit, manufacturing, medical, particle, size, source, vascular

## AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every five years but at least every 10 years. For a TIR, AAMI consults with a technical committee about five years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice, or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

**CAUTION NOTICE:** This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

or visit [www.aami.org](http://www.aami.org).

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633.

### *Published by*

Association for the Advancement of Medical Instrumentation  
4301 N. Fairfax Drive, Suite 301  
Arlington, VA 22203-1633  
[www.aami.org](http://www.aami.org)

© 2010 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, complete the reprint request form at [www.aami.org](http://www.aami.org) or contact AAMI at 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203. Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

**ISBN 1-57020-403-9**

# Contents

	Page
Glossary of equivalent standards .....	v
Committee representation .....	vii
Foreword .....	viii
Introduction .....	ix
1 Scope .....	1
1.1 General .....	1
1.2 Inclusion .....	1
1.3 Exclusion .....	1
2 Normative references .....	1
3 Definitions .....	2
4 Considerations for evaluating particulates .....	2
4.1 General .....	2
4.2 Identity .....	2
4.3 Size .....	2
4.4 Shape .....	2
4.5 Quantity .....	3
4.6 Test methods .....	3
5 Sources of particulates .....	3
5.1 Product design .....	3
5.1.1 Materials .....	3
5.1.2 Design and configuration .....	3
5.1.3 Manufacturability .....	3
5.1.4 Effect of packaging on device particulates .....	4
5.1.5 Effect of sterilization .....	4
5.1.6 Effect of shipping and distribution .....	4
5.1.7 Effect of shelf storage before use .....	5
5.1.8 Clinical use of device .....	5
5.2 Manufacturing .....	5
5.2.1 General .....	5
5.2.2 Particulates from materials .....	5
5.2.3 Particulates from manufacturing processes .....	6
5.2.4 Particulates from the environment .....	6
5.2.5 Particulates from personnel .....	7
5.2.6 Routine particulate monitoring and identification during manufacturing .....	7
6 Establishment of product particle limits .....	8
6.1 Allowable limits for particulates on a device .....	8
6.2 Clinical application of device .....	8
6.3 Visible particulate matter .....	8
6.4 Particle size ranges and particle count limits .....	8
6.5 Counting methods for fibrous particles .....	9
6.6 Alert limit .....	9
6.7 Action limit .....	9
6.8 Limits for raw materials and components .....	9
6.9 Product design .....	9



This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at (877) 249-8226 or visit [www.aami.org](http://www.aami.org).

7 Test methods for collection and analysis of particulate matter .....	9
7.1 General .....	9
7.2 Airborne particulate matter.....	10
7.2.1 Capturing airborne particulate matter for identification.....	10
7.2.2 Counting and sizing .....	10
7.3 Particulate matter in liquids .....	10
7.3.1 Capturing particulate matter in liquids for identification.....	10
7.3.2 Counting and sizing .....	10
7.3.3 Light obscuration counter (LOC) .....	11
7.3.4 Microscopic method .....	12
7.4 Characterization of particulate matter .....	14
7.4.1 Sample preparation.....	15
7.4.2 Organic analysis .....	15
7.4.3 Other inorganic methods.....	16
7.4.4 Morphological analysis.....	16
7.4.5 Other analytical techniques.....	17
8 Clinical significance of particulate matter .....	17
8.1 General .....	17
8.2 Therapeutic embolization .....	18
8.3 Microembolization in animals.....	19
8.4 Foreign body interactions.....	20
8.5 Conclusion.....	21
8.6 Cited references.....	21
8.7 Annotated bibliography .....	23



This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

<b>Annex</b>	For a complete copy of this AAMI document, contact AAMI at (877) 249-8226 or visit <a href="http://www.aami.org">www.aami.org</a> .....	39
A Summary of particulate technologies .....		39
<b>Bibliography</b> .....		44
<b>Tables</b>		
1 Disease type and effective size of microspheres .....		19
A.1 Particulate analysis technologies .....		39

## Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. For each International Standard that has been adopted by AAMI (and ANSI), the table below gives the corresponding U.S. designation and level of equivalency to the International Standard. NOTE—Documents are sorted by international designation. In the U.S. designation column, the code “(R)20xx” indicates the year the document was officially reaffirmed by AAMI. For example, ANSI/AAMI/ISO 10993-4:2002/(R)2009 indicates that 10993-4, originally approved and published in 2002, was reaffirmed without change in 2009.

Other normatively referenced International Standards may be under consideration for U.S. adoption by AAMI; therefore, this list should not be considered exhaustive.

International designation	U.S. designation	Equivalency
IEC 60601-1:2005 Technical Corrigendum 1 and 2	ANSI/AAMI ES60601-1:2005 and ANSI/AAMI ES60601-1:2005/A2:2010 ANSI/AAMI ES60601-1:2005/C1:2009 (amdt)	Major technical variations C1 Identical to Corrigendum 1 & 2
IEC 60601-1-2:2007	ANSI/AAMI/IEC 60601-1-2:2007	Identical
IEC 60601-2-2:2009	ANSI/AAMI/IEC 60601-2-2:2009	Identical
IEC 60601-2-4:201X <sup>1</sup>	ANSI/AAMI/IEC 60601-2-4:2010	Identical
IEC 60601-2-16:2008	ANSI/AAMI/IEC 60601-2-16:2008	Identical
IEC 60601-2-19:2009	ANSI/AAMI/IEC 60601-2-19:2009	Identical
IEC 60601-2-20:2009	ANSI/AAMI/IEC 60601-2-20:2009	Identical
IEC 60601-2-21:2009	ANSI/AAMI/IEC 60601-2-21:2009	Identical
IEC 60601-2-24:1998	ANSI/AAMI ID26:2004/(R)2009	Major technical variations
IEC 60601-2-47:2001	ANSI/AAMI EC38:2007	Major technical variations
IEC 60601-2-50:2009	ANSI/AAMI/IEC 60601-2-50:2009	Identical
IEC 80001-1:2010	ANSI/AAMI/IEC 80001-1:2010	Identical
IEC 80601-2-30:2009 and Technical Corrigendum 1	ANSI/AAMI/IEC 80601-2-30:2009 and ANSI/AAMI/IEC 80601-2-30:2009 C1:2009 (amdt) consolidated text	Identical (with inclusion) C1 Identical to Corrigendum 1
IEC 80601-2-58:2008	ANSI/AAMI/IEC 80601-2-58:2008	Identical
IEC/TR 60878:2009	ANSI/AAMI/IEC TIR60878:2003	Identical
IEC/TR 62296:2009	ANSI/AAMI/IEC TIR62296:2009	Identical
IEC 62304:2006	ANSI/AAMI/IEC 62304:2006	Identical
IEC/TR 62348:2006	ANSI/AAMI/IEC TIR62348:2006	Identical
IEC/TR 62354:2009	ANSI/AAMI/IEC TIR62354:2009	Identical
IEC 62366:2007	ANSI/AAMI/IEC 62377:2007	Identical
IEC/TR 80002-1:2009	ANSI/IEC/TR 80002-1:2009	Identical
ISO 5840:2005	ANSI/AAMI/ISO 5840:2005/(R)2010	Identical
ISO 7198:1998	ANSI/AAMI/ISO 7198:1998/2001/(R)2010	Identical
ISO 7199:2009	ANSI/AAMI/ISO 7199:2009	Identical
ISO 8637:2010	ANSI/AAMI/ISO 8637:2010	Identical
ISO 8638:2010	ANSI/AAMI/ISO 8638:2010	Identical
ISO 10993-1:2009	ANSI/AAMI/ISO 10993-1:2009	Identical
ISO 10993-2:2006	ANSI/AAMI/ISO 10993-2:2006	Identical
ISO 10993-3:2003	ANSI/AAMI/ISO 10993-3:2003/(R)2009	Identical
ISO 10993-4:2002 and Amendment 1:2006	ANSI/AAMI/ISO 10993-4:2002/(R)2009 and Amendment 1:2006/(R)2009	Identical
ISO 10993-5:2009	ANSI/AAMI/ISO 10993-5:2009	Identical
ISO 10993-6:2007	ANSI/AAMI/ISO 10993-6:2007	Identical
ISO 10993-7:2008	ANSI/AAMI/ISO 10993-7:2008	Identical
ISO 10993-9:2009	ANSI/AAMI/ISO 10993-9:2009	Identical
ISO 10993-10:2010	ANSI/AAMI/ISO 10993-10:2010	Identical
ISO 10993-11:2006	ANSI/AAMI/ISO 10993-11:2006	Identical
ISO 10993-12:2007	ANSI/AAMI/ISO 10993-12:2007	Identical
ISO 10993-13:2010	ANSI/AAMI/ISO 10993-13:2010	Identical
ISO 10993-14:2001	ANSI/AAMI/ISO 10993-14:2001/(R)2006	Identical
ISO 10993-15:2000	ANSI/AAMI/ISO 10993-15:2000/(R)2006	Identical
ISO 10993-16:2010	ANSI/AAMI/ISO 10993-16:2010	Identical
ISO 10993-17:2002	ANSI/AAMI/ISO 10993-17:2002/(R)2008	Identical
ISO 10993-18:2005	ANSI/AAMI BE83:2006	Major technical variations

<sup>1</sup> As of November 2010, in production

International designation	U.S. designation	Equivalency
ISO/TS 10993-19:2006	ANSI/AAMI/ISO TIR10993-19:2006	Identical
ISO/TS 10993-20:2006	ANSI/AAMI/ISO TIR10993-20:2006	Identical
ISO 11135-1:2007	ANSI/AAMI/ISO 11135-1:2007	Identical
ISO/TS 11135-2:2008	ANSI/AAMI/ISO TIR11135-2:2008	Identical
ISO 11137-1:2006	ANSI/AAMI/ISO 11137-1:2006/(R)2010	Identical
ISO 11137-2:2006 (2006-08-01 corrected version)	ANSI/AAMI/ISO 11137-2:2006	Identical
ISO 11137-3:2006	ANSI/AAMI/ISO 11137-3:2006/(R)2010	Identical
ISO 11138-1: 2006	ANSI/AAMI/ISO 11138-1:2006/(R)2010	Identical
ISO 11138-2: 2006	ANSI/AAMI/ISO 11138-2:2006/(R)2010	Identical
ISO 11138-3: 2006	ANSI/AAMI/ISO 11138-3:2006/(R)2010	Identical
ISO 11138-4: 2006	ANSI/AAMI/ISO 11138-4:2006/(R)2010	Identical
ISO 11138-5: 2006	ANSI/AAMI/ISO 11138-5:2006/(R)2010	Identical
ISO/TS 11139:2006	ANSI/AAMI/ISO 11139:2006	Identical
ISO 11140-1:2005	ANSI/AAMI/ISO 11140-1:2005/(R)2010	Identical
ISO 11140-3:2007	ANSI/AAMI/ISO 11140-3:2007	Identical
ISO 11140-4:2007	ANSI/AAMI/ISO 11140-4:2007	Identical
ISO 11140-5:2007	ANSI/AAMI/ISO 11140-5:2007	Identical
ISO 11607-1:2006	ANSI/AAMI/ISO 11607-1:2006	Identical
ISO 11607-2:2006	ANSI/AAMI/ISO 11607-2:2006	Identical
ISO 11663:2009	ANSI/AAMI/ISO 11663:2009	Identical
ISO 11737-1: 2006	ANSI/AAMI/ISO 11737-1:2006	Identical
ISO 11737-2:2009	ANSI/AAMI/ISO 11737-2:2009	Identical
ISO 13408-1:2008	ANSI/AAMI/ISO 13408-1:2008	Identical
ISO 13408-2:2003	ANSI/AAMI/ISO 13408-2:2003	Identical
ISO 13408-3:2006	ANSI/AAMI/ISO 13408-3:2006	Identical
ISO 13408-4:2005	ANSI/AAMI/ISO 13408-4:2005	Identical
ISO 13408-5:2006	ANSI/AAMI/ISO 13408-5:2006	Identical
ISO 13408-6:2006	ANSI/AAMI/ISO 13408-6:2006	Identical
ISO 13485:2003	ANSI/AAMI/ISO 13485:2003/(R)2009	Identical
ISO 14155-1:2003	ANSI/AAMI/ISO 14155-1:2003/(R)2008	Identical
ISO 14155-2:2003	ANSI/AAMI/ISO 14155-2:2003/(R)2008	Identical
ISO 14160:1998	ANSI/AAMI/ISO 14160:1998/(R)2008	Identical
ISO 14161:2009	ANSI/AAMI/ISO 14161:2009	Identical
ISO 14708-3:2008	ANSI/AAMI/ISO 14708-3:2008	Identical
ISO 14708-4:2008	ANSI/AAMI/ISO 14708-4:2008	Identical
ISO 14708-5:2010	ANSI/AAMI /ISO 14708-5:2010	Identical
ISO 14937:2009	ANSI/AAMI/ISO 14937:2009	Identical
ISO/TR 14969:2004	ANSI/AAMI/ISO TIR14969:2004	Identical
ISO 14971:2007	ANSI/AAMI/ISO 14971:2007/(R)2010	Identical
ISO 15223-1:2007 and A1:2008	ANSI/AAMI/ISO 15223-1:2007 and A1:2008	Identical
ISO 15223-2:2010	ANSI/AAMI/ISO 15223-2:2010	Identical
ISO 15225:2010	ANSI/AAMI/ISO 15225:2010	Identical
ISO 15674:2009	ANSI/AAMI/ISO 15674:2009	Identical
ISO 15675:2009	ANSI/AAMI/ISO 15675:2009	Identical
ISO 15882:2008	ANSI/AAMI/ISO 15882:2008	Identical
ISO 15883-1:2006	ANSI/AAMI ST15883-1:2009	Major technical variations
ISO/TR 16142:2006	ANSI/AAMI/ISO TIR16142:2005	Identical
ISO 17664:2004	ANSI/AAMI ST81:2004	Major technical variations
ISO 17665-1:2006	ANSI/AAMI/ISO 17665-1:2006	Identical (with inclusions)
ISO/TS 17665-2:2009	ANSI/AAMI/ISO TIR17665-2:2009	Identical
ISO 18472:2006	ANSI/AAMI/ISO 18472:2006	Identical
ISO/TS 19218:2005	ANSI/AAMI/ISO 19218:2005	Identical
ISO 22442-1:2007	ANSI/AAMI/ISO 22442-1:2007	Identical
ISO 22442-2:2007	ANSI/AAMI/ISO 22442-2:2007	Identical
ISO 22442-3:2007	ANSI/AAMI/ISO 22442-3:2007	Identical
ISO 25539-1:2003 and A1:2005	ANSI/AAMI/ISO 25539-1:2003/(R)2009 and A1:2005/(R)2009	Identical
ISO 25539-2:2008	ANSI/AAMI/ISO 25539-2:2008	Identical
ISO 27186:2010	ANSI/AAMI/ISO 27186:2010	Identical
ISO 81060-1:2007	ANSI/AAMI/ISO 81060-1:2007	Identical
ISO 81060-2:2009	ANSI/AAMI/ISO 81060-2:2009	Identical



## Committee representation

### Association for the Advancement of Medical Instrumentation

#### Medical Device Particulates Committee

This technical information report (TIR) was developed by the AAMI Medical Device Particulates Committee. Approval of the TIR does not necessarily imply that all committee members voted for its approval.

At the time this document was published, the **AAMI Medical Device Particulates Committee** had the following members:

- Co-chairs:* Silvia Garcia-Tunon, MS MSEM  
Christopher Siedlecki, PhD
- Members:* Susanne Anderson, NAMSA  
Kirk Ashline, Baxter Healthcare Corporation  
Brian Choules, PhD, MED Institute, Inc.  
James Conti, PhD, Dynatek Dalta Scientific Instruments  
Paul Fioriti, Covidien  
Silvia Garcia-Tunon, MS MSEM, Cordis/Johnson & Johnson  
Doug Harbrecht, Boston Scientific Corporation  
Stephen Hilbert, PhD MD, Children's Mercy Hospital/Ward Family Center for Congenital Heart Disease  
Nelson, Huldin, PMP, Hospira Worldwide, Inc.  
Greg LoStracco, WL Gore & Associates, Inc.  
Ronald Lulich, 3M Healthcare  
Tonya Morris, BS, Nelson Laboratories, Inc.  
Kim-Lien Nguyen-Ehrenreich, MS, Abbott Laboratories  
Lori Osterby, St. Jude Medical  
Dinesh Patwardhan, U.S. Food and Drug Administration/Center for Devices and Radiological Health  
Matthew Russell, Cook Inc.  
Michael Scholla, DuPont Nonwovens  
Christopher Siedlecki, PhD, Pennsylvania State University College of Medicine  
Mark Slaughter, MD, Oak Lawn, IL  
Sean Stucke, SurModics, Inc.  
John Toigo, Terumo Heart, Inc.  
Rick Tyson, WuXi Apptec
- Alternates:* Carol Arnett, WuXi AppTec  
Stephanie Del Paine, PhD, MED Institute, Inc.  
Shawn Fuller, SurModics, Inc.  
Ralph Makinen, Boston Scientific Corporation  
David McGoldrick, BS, Abbott Laboratories  
Hina Pinto, U.S. Food and Drug Administration/Center for Devices and Radiological Health  
Deanna Porter, St. Jude Medical  
Elaine Strope, PhD, Dynatek Dalta Scientific Instruments

---

NOTE—Participation by federal agency representatives in the development of this technical information report does not constitute endorsement by the federal government or any of its agencies.

---

## Foreword

This technical information report was developed by the AAMI Medical Device Particulates Committee. The objective is to provide technical information that will assist medical device manufacturers in determining acceptable levels of medical device particulates on products used to deliver or implant into the vasculature, or both.

As used within the context of this document, “should” indicates that among several possibilities one is recommended as particularly suitable without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action should be avoided but not prohibited. “May” is used to indicate that a course of action is permissible within the limits of the TIR. “Can” is used as a statement of possibility and capability. “Must” is used only to describe “unavoidable” situations, including those mandated by government regulations.

Suggestions for improving this technical information report are invited. Comments and suggested revisions should be sent to Technical Programs, AAMI, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633.



---

NOTE—This foreword does not contain provisions of the AAMI TIR42, *Evaluation of particulates associated with vascular medical devices* (AAMI TIR42:2010), but it does provide important information about the development and intended use of the document.

---

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document,  
contact AAMI at (877) 249-8226  
or visit [www.aami.org](http://www.aami.org).

## Introduction

This technical information report (TIR) addresses the particulate matter on implantable medical devices and on accessory devices used in the vascular system during the delivery and implantation or exposure and removal of such devices. Unintentional particulate matter on medical devices may be a quality control issue because of the manufacturing environment or a device design-related issue. Sources of particles in the manufacturing environment may include glove powders, lint and other fibers, paper particles, packaging materials, paint particles, and many other materials. Shedding of particles during use is a characteristic of medical devices that can be addressed in product development. Particles consisting of device materials may arise because of friction, abrasion, dissolution, or intentional degradation, and can have significant effects on patient outcome.



This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document,  
contact AAMI at (877) 249-8226  
or visit [www.aami.org](http://www.aami.org).



This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document,  
contact AAMI at (877) 249-8226  
or visit [www.aami.org](http://www.aami.org).

# Evaluation of particulates associated with vascular medical devices

## 1 Scope

### 1.1 General

This technical information report (TIR) addresses vascular exposure to particles arising from the manufacturing environment for vascular medical devices and from the use of vascular medical devices. It is intended to assist vascular medical device manufacturers in determining the source of particulates, establishing product particulate limits, defining appropriate test methods, and assessing the clinical relevance of particulate contamination.

### 1.2 Inclusion

This document specifically includes particles generated as a result of manufacturing, packaging, and acute application of a vascular medical device or any accessories bundled with the device used for insertion and implementation of the product. In the absence of bundled accessories, this TIR recommends testing the device in conjunction with commonly used or suggested accessories.

### 1.3 Exclusion

This TIR does not address particulates that arise from degradation or wear of the device, whether these particles be either deliberate or nondeliberate. This document specifically excludes particles arising from the operating room or clinical environment in which the device might be implemented. Furthermore, the scope of this document specifically excludes visible particles that would normally be found using standard visual inspection techniques that are part of the manufacturing process. This document does not address patient-generated particles that might be produced before, during, or following implementation of the device. Liquids are not considered to be particles in the context of this document. In addition, therapeutic particles intentionally delivered are not included in this document, although characterization of these particles using methods described in this document may be appropriate.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. The latest edition of the referenced documents (including any amendments) applies.

ANSI/AAMI/ISO 10993-1, *Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process*

ASTM F24, *Standard test method for measuring and counting particulate contamination on surfaces*

European Directorate for the Quality of Medicines & Healthcare (EDQM), *European pharmacopeia*

Institute of Environmental Sciences and Technology (IEST): IEST-STD-CC1246D, *Product cleanliness levels and contamination control program*

ISO 14644 (series), *Cleanrooms and associated controlled environments*

U.S. Pharmacopoeia (USP) <788>, *Particulate matter in injections*